Rx Only


FOR INTRAOCULAR USE AS A SURGICAL AID IN ANTERIOR SEGMENT SURGERY ONLY.

DESCRIPTION

Healon® EndoCoat Ophthalmic Viscosurgical Device (OVD) is a sterile non-pyrogenic solution of highly purified sodium hyaluronate with rheologically dispersive properties. Healon® EndoCoat OVD contains 30 mg/ml of sodium hyaluronate. The concentration is adjusted to yield a viscosity of approximately 50,000 centipoise when dissolved in a physiological buffered salt solution. The solution has an osmolality of approximately 320 milliosmoles/kg.

CHARACTERISTICS

Sodium hyaluronate is a linear polysaccharide composed of repeating disaccharides of sodium glucuronate and N-acetylglucosamine found throughout the tissues of the body with high concentrations in the vitreous humor, synovial fluid, and umbilical cord. It has a role in regulating the interactions between adjoining tissues. Sodium hyaluronate can also act as a viscoelastic space filler, maintaining a



separation between tissues. Sodium hyaluronate does not interfere with the normal wound healing process. Sodium hyaluronate is also present in the capsular material of certain bacteria. Those bacteria may be cultured by a fermentation process to yield sodium hyaluronate. Sodium hyaluronate extracted and purified from different sources can have different molecular weights but has the same molecular structure. The sodium hyaluronate in Healon® EndoCoat OVD is a highly purified extract of a bacterial fermentation and is tolerated well in the eye. It has an average molecular weight of approximately 800,000 Daltons is non-antigenic (1, 2, 3), and non-pyrogenic.

INDICATIONS

Healon® EndoCoat OVD is an ophthalmic viscoelastic containing 3% sodium hyaluronate indicated for use as a surgical aid in patients undergoing ophthalmic anterior segment procedures including:

- Cataract surgery with an intraocular lens
- Cataract surgery without an intraocular lens
- Secondary intraocular lens implantation

Healon® EndoCoat maintains a deep chamber during anterior segment surgery, aids in tissue manipulation during surgery, enhances visualization during the surgical procedure and protects the corneal endothelium and other ocular tissue. The viscoelasticity of the solution maintains the normal position of the vitreous face and prevents formation of a flat chamber during surgery. It may also be used to coat intraocular lenses and insertion instruments prior to intraocular lens implantation.

CONTRAINDICATIONS

At present, there are no contraindications to the use of Healon® EndoCoat OVD when used as recommended.

WARNINGS

The Healon® EndoCoat Delivery system is not designed or intended to be attached to instruments other than the one provided with the product. Failure to follow the "Directions for Use" may result in cannula detachment.

Mixing of quaternary ammonium salts, such as benzalkonium chloride, with sodium hyaluronate results in the formation of a precipitate. The eye should not be irrigated with any solution containing benzalkonium chloride if Healon® EndoCoat is to be used during surgery.

DIRECTIONS FOR USE

Cataract surgery and intraocular lens (IOL) implantation: Inject Healon® EndoCoat OVD slowly through the cannula into the anterior chamber. The use of Healon® EndoCoat OVD is most effective when the injection is made before phacoemulsification removal of the cataract and before insertion of the IOL. Healon® EndoCoat OVD can also be applied to the IOL before placement. During the procedure, more Healon® EndoCoat OVD can be infused for anterior chamber maintenance or to replace viscoelastic lost during surgery. At the end of the surgical procedure, it is required that Healon® EndoCoat OVD be completely removed from the eye by thoroughly irrigating and aspirating with a sterile irrigation solution. Due to the adherent nature of a dispersive viscoelastic more time and care may be required to remove the viscoelastic completely from the eye.

PRECAUTIONS

CAUTION: Injection of viscoelastics creates pressure in the syringe. To prevent

expulsion of the cannula into the eye, ensure that the cannula is securely

attached to the fitting on the syringe. Use of the cannula guard is

recommended.

CAUTION: The cannula should be fastened securely to the syringe; however, over

tightening may cause the hub to weaken and possibly detach from the syringe. Extrusion of a test droplet is recommended prior to entering the

eye, and excessive force on the plunger should be avoided.

CAUTION: Do not reuse the cannula. This could release particulate matter. Product

and cannula are for single use only. Re-use may cause eye inflammation.

CAUTION: The potential for early and short-term postoperative intraocular pressure

(IOP) spikes exists with dispersive OVDs, which potentially require more time and care to remove from the eye. Therefore, it is recommended that Healon Endocoat OVD be removed from the eye completely by irrigating and aspirating with sterile irrigation solution to reduce the risk of early

postoperative IOP spikes.

Observe the usual precautions taken during anterior segment surgery.

Pre-existing glaucoma, the surgery itself, or retained viscoelastic (particularly in patients with compromised trabecular meshwork) can cause increased intraocular pressure after the procedure (4).

The following precautions should be carefully considered:

- The intraocular pressure of postoperative patients should be carefully monitored, particularly in the early post operative period.
- Do not use excessive amounts of Healon® EndoCoat OVD.
- Remove Healon® EndoCoat OVD completely from the anterior chamber at the end of the procedure.
- Corrective therapy should be initiated if the postoperative intraocular pressure rises above safe levels.
- For intraocular use only
- Store at 2-25°C (36-77°F)
- Protect from freezing.
- Protect from light.
- Use only if solution is clear
- Avoid trapping air bubbles
- Contents are sterile when the package is sealed and undamaged.
- Use aseptic technique.
- Do not use in cases of known hypersensitivity to any of the ingredients of this product.
- See product expiration date.

Healon® EndoCoat OVD does not require refrigeration. If refrigerated, Healon® EndoCoat OVD should be allowed to attain room temperature prior to use.

There have been isolated reports of diffuse particulates or haziness appearing after injection of viscoelastics into the eye. While such reports are infrequent and seldom associated with any effects on ocular tissue, the physician should be aware of this occurrence. If observed, the viscoelastic should be removed by irrigation and/or aspiration.

Healon® EndoCoat OVD is derived from microbial fermentation by a purified proprietary process. Although precautions have been taken to make this device protein-free, it may contain trace amounts of protein. The physician should be aware of the potential allergic risks such as postoperative inflammation that can occur with injection of biological materials.

Adverse Reactions

Because sodium hyaluronate is a polysaccharide present in many tissues of the body, it is extremely well tolerated in human eyes. There have been reports of transient postoperative ocular inflammation (oral and/or topical steroid treatments were administered) (5) and transient postoperative increases in intraocular pressure (4) during clinical trials with viscoelastics.

In addition to the above, the following adverse reactions have been reported following the use of sodium hyaluronate in intraocular surgery: inflammation, corneal edema, increased intraocular pressure, secondary glaucoma and corneal decompensation.

All of the adverse reactions described above are potential adverse reactions for Healon® EndoCoat OVD.

Refer to the Clinical Trial Section for more details regarding the adverse reactions (i.e., adverse events) that occurred in our study and the incidence rates.

Clinical Trial

Healon® EndoCoat OVD in the 2.25mL configuration was compared to Viscoat® in a randomized, double-masked, multi-center clinical trial between September 10, 2009, and August 11, 2010. Four hundred subjects (200 Healon® EndoCoat and 200 Viscoat®) with cataracts who were otherwise healthy, nonglaucomatous patients who did not receive prophylactic IOP lowering medications were evaluated for safety and effectiveness.

This study was conducted to evaluate the safety and effectiveness of an investigative dispersive ophthalmic viscoelastic under normal use conditions during phacoemulsification cataract surgery and intraocular lens placement as compared to a currently marketed dispersive OVD.

Safety was assessed by two primary endpoints: Cumulative rate of Intraocular Pressure (IOP) spikes ≥ 30 mm Hg and mean percent change in corneal endothelial cell count (ECC) between the preoperative and three months postoperative visits.

Following surgery, intraocular pressure (IOP) was measured at 6 hours, 24 hours, 7 days, 1 month and 3 months. Study results demonstrate that for the percentage of subjects with an IOP spike defined as IOP \geq 30 mm Hg (subjects did not receive IOP lowering medications at the time of surgery), Healon® Endocoat is non-inferior to Viscoat® with regard to the cumulative percentage of subjects with IOP \geq 30 mm Hg during the study (p=0.0003, δ =0.13, 90% confidence interval (-1.74, 7.72)).

Operative complications were reported in the safety population for 3% (6/200) of subjects in the Healon® EndoCoat group and 8% (16/200) in the Viscoat® group.

Operative Complications Safety Population

Operative Parameters		Hea Endo N =	Coat	Vísc N =	
		n	%	n	%
Surgical Complications	None	194	97.0	184	92.0
	Detached Descemet's membrane	2	1.0	2	1.0
	Iris damage	0	0.0	1	0.5
	Endothelial touch	0	0.0	2	1.0
	Capsule rupture/tear	2	1.0	6*	3.0
	Vitreous bulge or loss	1	0.5	2* [†]	1.0
	Zonular rupture	0	0.0	1	0.5
	Corneal abrasion	1	0.5	0	0.0
	IOL exchange	0	0.0	2	1.0

	Healon EndoCoat N = 200		Visc N = :	
Iris retraction	0	0.0	1	0.5

Safety Population – analysis population consisting of all subjects exposed to an OVD.

Percent IOP ≥30 mmHg by Viscoelastic Group, Visit and Cumulative Intent To Treat (ITT) Population*

Visit	VISCOELASTIC GROUP	N		P spikes (%)/ of IOP Spikes)
6 Hours	Healon® EndoCoat	199	7.5	(15)
	Viscoat	201	6.1	(12)
	Difference	-	1.4	
1 Day	Healon® EndoCoat	199	2.5	(5)
	Viscoat	201	2.0	(4)
	Difference	-	0.5	
1 Week	Healon® EndoCoat	199	1.0	(2)
	Viscoat	201	1.0	(2)
	Difference	-	0	
1 Month	Healon® EndoCoat	199	0	
	Viscoat	201	0	
	Difference	-	0	
3 Month	Healon® EndoCoat	199	0	
	Viscoat	201	., 0	
	Difference	-	0	
Cumulative	Healon® EndoCoat	199	10.6	(21)
	Viscoat	201	7.6	(15)
	Difference	-	3.0	(-1.74, 7.72)**

Intent-to-Treat (ITT) primary analysis population where missing values were imputed

Missing values are imputed by MCMC multiple imputation. Therefore, percent IOP>30 mmHG rather the

Post-Operative IOP Spike Rate Over Time Safety Population* Healon Endocoat

IOP	6 Hc			Day 200	1 W N =					onths = 200
	n	%	n	%	n	%	n	%	n	%
< 30 mmHg	185	92.5	195	97.5	197	98.9	200	100.0	200	100.0
≥30-39 mmHg	11	5.5	4**	2.0	2**	1.0	0	0.0	0	0.0

^{*}One subject had both a capsular rupture and vitreous bulge. One subject had both a capsulotomy and a vitrectomy.

[†]One subject had both a capsulotomy and a vitrectomy.

Missing values are imputed by MCMC multiple imputation. Therefore, percent IOP≥30 mmHG rather than number of IOP≥30 mmHG is reported.

^{90%} Confidence interval. Healon® EndoCoat is statistically significantly non-inferior to Viscoat (p=0.0003)

IOP	6 Hc			Day ∶200	1 W N =			onth 200	3 Mo N =	
	n	%	n	%	n	%	n	%	n	%
≥40-49 mmHg	2	1.0	1	0.5	0	0.0	0	0.0	0	0.0
≥ 50-60 mmHg	2	1.0	0	0.0	0	0.0	0	0,0	0	0.0

^{*} In the absence of IOP lowering medication administered at the time of surgery

Post -Operative IOP Spike Rate Over Time Safety Population*

Viscoat

IOP		ours 198		Day 200	1 W. N ≈			Month = 198		onths = 199
	n	%	n	%	n	%	n	%	n	%
< 30 mmHg	186‡	93.9	196‡	98.0	197	99.0	198	100.0	199	100.0
≥30-39 mmHg	10** [†]	5.1	2 [†]	1.0	1‡	0.5	0	0.0	0	0.0
≥40-49 mmHg	2	1.0	2**	1.0	1**	0.5	0	0.0	0	0.0
≥ 50-60 mmHg	0	0.0	0	0.0	0	0.0	0	0.0	0	0.0

^{*} In the absence of IOP lowering medication administered at the time of surgery


^{**} One subject experienced an IOP spike one day postoperatively of 38 mmHg and an IOP of 31 mm Hg at one week postoperatively off of IOP lowering medications).

^{**} One subject experienced an IOP of 39 mmHg at six hours postoperatively. Although IOP lowering medications and/or treatments were administered, the IOP rose to 41 mmHg at one day postoperatively. The IOP measured 46 mmHg at one week postoperatively off of IOP lowering medications;

One subject had an IOP of 35 mmHg six hours postoperatively. After IOP lowering medications and/or treatments were administered the IOP measured 32 mmHg at one day postoperatively.

[‡] One subject had an IOP of 38 mmHg at 1 week postoperatively, which had been preceded by unmedicated IOPs of 28 mmHg at 6 hours postoperatively and 26 mmHg at 1 day postoperatively.

Healon Endocoat demonstrated non-inferiority compared to Viscoat in corneal endothelial cell count (ECC) from preoperative to three months postoperative (p<0.0001, 1-sided t-test, δ = 5%). The observed mean percent changes in ECC from preoperative to three months postoperative for Healon® EndoCoat and Viscoat® were -4.7% and -7.0%, respectively with a 2.3 percentage point difference (90% CI: (0.23%, 4.33%)).



The distribution of postoperative medical findings/observations was similar between the two study groups and within the range of what would typically be reported. In the early postoperative period, inflammatory cells in the anterior chamber were the most reported form of inflammation for both viscoelastic groups. Reports of inflammatory cells diminished over time to minimal levels by the one-month visit in both viscoelastic groups. Early postoperative incidence rates of corneal, epithelial and stromal edema were low with similar results in both groups, diminishing over time. For other general slit-lamp findings, the majority of subjects in both groups were reported as "none" at all postoperative visits.

Clinical Trial Adverse Events Safety Population

Thirty nine subjects experienced adverse events in the study. None of the adverse events were considered unanticipated. Ninety-two percent of the adverse events were IOP \geq 30 mmHg; incidence of IOP \geq 30 mmHg occurred at a rate of 10.5% in the Healon® EndoCoat group, and 7.5% in the Viscoat® group. The three adverse events not related to IOP \geq 30 mmHg include: one subject in the Healon® EndoCoat group who developed cystoid macular edema (CME) requiring treatment and two subjects in the Viscoat group; one who underwent a lens explant in the study eye due to a shorn haptic and another who had an intraocular foreign body removed from the study eye. None of these three events was considered by the investigators to be related to the viscoelastic used.

ADVERSE EVENT		lon [®] Coat 200	Viscoat [®] N = 200		
	n	%	n	%	
Elevated IOP ≥ 30 mmHg	21*	10. 5	15*	7.5	
IOL Exchange	0	0.0	. 1	0.5	
Treatment of CME	1	0.5	0	0.0	
Removal of Foreign Body**	0	0.0	1	0.5	
Total Subjects Experiencing Adverse Events	22		17		

^{*} One subject in each group experienced two separate incidences of IOP ≥ 30 mmHg.

Although not included in the primary outcome parameters, the surgical time spent on OVD removal was recorded during the study. A subjective response was requested from surgeons at the end of the case regarding the ease of removal of the viscoelastic (choices were: easy, average, difficult, or hard).

Safety Population
Operative Parameters - Removal of Viscoelastic agent

Viscoelastic Re	End	alon ocoat 200	Viscoat N = 200		
Viscoelastic removal time (seconds)	(149.1		33.7
	SD 37.92		.92	35.21	
	Median 135.0		121.0		
	Min	(30	60	
	Max	3	00	. 454	
Ease of viscoelastic removal (no. of cases)	Easy	7	3.5%	9	4.5%
	Average	126	63%	142	71%
	Difficult	66	33%	49	24.5%
	Very Difficult	1	0.5%	0	0%

^{**} Intraocular foreign body was noted at the one month postoperative visit

HOW SUPPLIED

Healon® EndoCoat OVD is a sterile, non-pyrogenic preparation supplied in a disposable single-use glass syringe, delivering 0.85 ml of a solution of sodium hyaluronate in a physiological buffered salt solution.

A sterile, single-use 25-gauge, disposable, bent, blunt-tip thin-wall cannula and cannula guard are provided within the package. The cannula sheath should be used to firmly attach the cannula to the syringe. Contents of unopened and undamaged blister package are sterile. Do not use if package is opened or damaged.

Contents

Each ml of Healon® EndoCoat OVD contains:

Ingredient	Contents per mi
sodium hyaluronate	30.00 mg
sodium chloride	5.00 mg
potassium chloride	0.56 mg
calcium chloride	0.36 mg
magnesium chloride	0.22 mg
sodium acetate	2.92 mg
sodium citrate	1.28 mg
sodium phosphate dibasic	0.42 mg
sodium phosphate monobasic	0.06 mg
water for injection	as required

Healon® EndoCoat OVD exhibits an osmolality of approximately 320 mOsm/kg and a pH of 6.8 - 7.6. Healon® EndoCoat OVD is filter sterilized and aseptically filled. The packaged product is secondarily sterilized using ethylene oxide.

REFERENCES

- 1. Richter W. Non-immunogenicity of purified hyaluronic acid preparations tested by passive cutaneous anaphylaxis. Int Arch Allergy 1974; 47:211.
- 2. Richter, W. Ryde EM, Zetterstrom EO. Non-immunogenicity of a purified sodium hyaluronate preparation in man. Int Arch Appl Immunol 1979; 59:45.
- 3. Lifecore Biomedical document of antigenicity studies.
- 4. Miller D, Stegmann R. The use of Healon in intraocular lens implantation. Int Ophthalmol clinics 1982; 22:177.
- 5. Pruett RC, Schepens CL, Swann DA. Hyaluronic acid vitreous substitute. A six year clinical evaluation. Arch Ophthalmol 1979; 87:699.

Caution: Federal (USA) law restricts this device to sale, distribution, or use by or on the order of a physician.

SYMBOLS USED ON STERILE PACKAGING

SYMBOLS	ENGLISH
	Consult Instructions For Use
<u> </u>	Caution
<u></u>	Do Not Use if Package is Damaged
8	Do Not Reuse
STERILE A	Sterile using Aseptic Processing Techniques
STERILEEO	Packaging and Cannula Sterilized using Ethylene Oxide
REF	Catalogue Number
LOT	Lot Number
2	USE BY (YYYY-MM: year-month)
类	Protect From Light
8	Protect From Freezing
X	Temperature Limitation
Ш	Manufacturer
EC REP	Authorised Representative in the European Community

Healon and Healon® EndoCoat are trademarks owned by or licensed to Abbott Laboratories, its subsidiaries or affiliates.



Abbott Medical Optics Inc. 1700 E. St. Andrew Place Santa Ana, CA 92705 USA www.amo-inc.com

1-877-AMO-4-LIFE



EC REP

AMO Ireland Block B Liffey Valley Office Campus Quarryvale, Co. Dublin, Ireland

Product of USA


©2012 Abbott Medical Optics Inc.



Healon® EndoCoat Opthalmic Viscosurgical Device

Graphics

Remove plastic tip cap from syringe tip.




7 Thread the cannula onto the syringe and confirm that it is firmly seated.




Remove the plastic sheath from the cannula by pulling it away in a straight motion.

Ensure that cannula remains fully seated to syringe.



4 Guide the cannula needle through the opening of the cannula guard provided until the cannula guard is fully seated against the cannula hub.

Click the cannula guard in place around the syringe.



Check proper function by holding the syringe barrel and gently depressing the plunger rod until the OVD appears at the cannula tip.

